

EC DECLARATION OF CONFORMITY

Name and address of the manufacturer: Foshan COXO Medical Instrument Co., Ltd.
No. 17, Guangming Ave., New Light Source Industrial Base, Nanhai National High-tech Zone, Foshan 528226, Guangdong P.R. China
SRN: CN-MF-000001682

EC Representative: Lotus NL B.V.
Koningin Julianaplein 10, 1e Verd, 2595AA, The Hague, Netherlands.
E-mail: peter@lotusnl.com
SRN: NL-AR-000000121

Trademarks: **COXO, YSDENT**

We declare under our sole responsibility that

the medical device: **Product Name:** Dental Scaler and Air Polisher
Model: PT Master
Basic UDI-DI: 697426789372G
GMDN: 45403

MDA Code: 0311 Active non-implantable dental devices

Intended Use: This product is intended only for use in the field of dentistry, for surgery to tooth surface cleaning, removal of foreign substances, root canal obturation, enhancing adhesion of fixed restorations, incision or removal of periodontal tissue; The device is only used in dental clinics or hospitals.

of class: IIa, according to Rule 9, Annex VIII, Regulation (EU) 2017/745

meets the provisions of Medical Device Regulation (EU) 2017/745. The declaration is valid in connection with the "final inspection report" of the device.

Conformity assessment procedure: Annex IX (Full QMS), Regulation (EU) 2017/745

Registration No.: **HZ 2182788-1**

Notified Body: **TÜV Rheinland LGA Products GmbH**
Tillystraße 2
90431 Nürnberg
Deutschland
CE 0197

(FoShan), PR China 2024-10-12

Place, date

